
Midmark M11 - v1.0.0
Total Cycles: 2017

Name: RITTER 1
Sterilizer ID: V2178198
06/09/2022 12:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.2 F 0.0
2:00 172.7 F 3.5
4:00 201.8 F 7.9
6:00 225.2 F 13.5
8:00 224.7 F 7.8
10:00 237.9 F 13.0
12:00 237.3 F 9.9
14:00 236.5 F 8.4
16:00 248.7 F 14.7
18:00 261.0 F 21.6
20:00 272.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.7 F 29.2
2:00 271.7 F 28.4
3:00 271.6 F 28.9
4:00 273.0 F 29.4

Min 271.0 F 28.2
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 216.6 F 0.0
10:00 207.1 F 0.0
15:00 242.7 F 0.0
20:00 229.6 F 0.0
25:00 235.3 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 20:31
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:15
CYCLE COMPLETE

06/09/2022 01:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____