
Midmark M11 - v1.0.0
Total Cycles: 2180

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2022 02:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.0 F 0.0
2:00 202.4 F 2.8
4:00 211.8 F 8.4
6:00 226.5 F 9.4
8:00 235.7 F 10.2
10:00 234.8 F 7.8
12:00 244.1 F 12.1
14:00 245.0 F 12.7
16:00 257.9 F 19.5
18:00 268.9 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.1 F 28.6
2:00 271.9 F 28.8
3:00 271.7 F 28.4
4:00 272.3 F 29.0

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 231.2 F 0.0
10:00 219.5 F 0.0
15:00 207.0 F 0.0
20:00 222.0 F 0.0
25:00 220.1 F 0.0
Dry Cycle Aborted

FILLING: 1:27
HEATING: 19:03
STERILIZING: 4:00
VENTING: 3:14
DRYING: 27:16
TOTAL CYCLE: 00:55:01

07/26/2022 03:00 PM
ACCEPTED: X
REJECTED:
Operator: U02