
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000209
7:01 22/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:05
STERILIZING 4:19
136.0 C 230kPa 4:19
136.3 C 232kPa 4:49
136.0 C 229kPa 5:19
136.2 C 232kPa 5:49
136.1 C 230kPa 6:19
136.3 C 231kPa 6:49
136.3 C 230kPa 7:19
136.4 C 232kPa 7:49
136.4 C 232kPa 7:49
VENTING 7:49
AIR DRYING 8:42
CYCLE COMPLETE 10:24

Digital Signature #
4281BDFEFC7B1D2A
