
Midmark M11 - v1.0.0
Total Cycles: 1009

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2021 09:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.5 F 0.0
2:00 181.7 F 3.3
4:00 197.0 F 7.9
6:00 214.2 F 13.2
8:00 223.5 F 9.1
10:00 236.6 F 9.7
12:00 241.6 F 12.1
14:00 240.2 F 10.5
16:00 252.7 F 16.9
18:00 263.9 F 23.5
20:00 271.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.8 F 29.2
2:00 271.2 F 28.5
3:00 272.6 F 29.3
4:00 272.1 F 28.9

Min 271.0 F 28.5
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 235.0 F 0.0
10:00 228.9 F 0.0
15:00 226.4 F 0.0
20:00 222.9 F 0.0
25:00 227.6 F 0.0
30:00 200.5 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 20:06
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

07/14/2021 10:44 AM
ACCEPTED: X
REJECTED:
Operator: U01