
Midmark M11 - v1.0.5
Total Cycles: 3357

Name: RITTER2
Sterilizer ID: V2313228
02/05/2024 04:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.5 F 0.0
2:00 127.4 F 0.5
4:00 136.5 F 2.7
6:00 168.5 F 6.2
8:00 186.0 F 10.4
10:00 213.2 F 6.0
12:00 221.9 F 8.4
14:00 233.7 F 13.0
16:00 234.9 F 10.1
18:00 236.3 F 9.1
20:00 247.4 F 14.6
22:00 259.0 F 20.9
24:00 269.4 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.7 F 29.6
2:00 271.3 F 29.3
3:00 272.6 F 29.7
4:00 271.5 F 29.1

Min 271.2 F 29.0
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.1 F 0.0
5:00 209.1 F 0.0
10:00 214.5 F 0.0
15:00 210.6 F 0.0
20:00 208.2 F 0.0
25:00 210.0 F 0.0
30:00 211.2 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 25:01
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:03:39
CYCLE COMPLETE

02/05/2024 05:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____