

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4315

Name: RITTER 1  
Sterilizer ID: V2178198  
02/21/2024 12:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.2 F 0.0  
2:00 153.7 F 2.1  
4:00 158.9 F 5.2  
6:00 180.1 F 8.8  
8:00 198.1 F 12.8  
10:00 217.0 F 6.0  
12:00 227.9 F 10.0  
14:00 229.6 F 8.9  
16:00 238.0 F 10.6  
18:00 235.1 F 8.2  
20:00 247.0 F 14.0  
22:00 257.8 F 19.7  
24:00 267.9 F 25.9  
26:00 272.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.0  
1:00 272.5 F 29.2  
2:00 272.7 F 29.5  
3:00 271.8 F 28.9  
4:00 272.7 F 29.3

Min 271.8 F 28.9  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.9 F 0.0  
5:00 218.5 F 0.0  
10:00 227.2 F 0.0  
15:00 226.8 F 0.0  
20:00 231.9 F 0.0  
25:00 231.8 F 0.0  
30:00 226.2 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 26:12  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:05:08  
CYCLE COMPLETE

02/21/2024 01:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_