
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003852
12:28 21/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:48
STERILIZING 3:52
136.4 C 223kPa 3:52
136.2 C 223kPa 4:22
136.4 C 224kPa 4:52
136.3 C 223kPa 5:22
136.3 C 223kPa 5:52
136.2 C 224kPa 6:22
136.3 C 219kPa 6:52
136.3 C 223kPa 7:22
136.3 C 223kPa 7:22
VENTING 7:22
AIR DRYING 8:12
CYCLE COMPLETE 9:51

Digital Signature #
56EC0BC9C0B3E4C4
