
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001283
15:42 03/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:24
STERILIZING 4:31
136.0 C 220kPa 4:31
136.4 C 225kPa 5:01
136.2 C 221kPa 5:31
136.3 C 224kPa 6:01
136.2 C 222kPa 6:31
136.1 C 222kPa 7:01
136.3 C 224kPa 7:31
136.4 C 225kPa 8:01
136.4 C 225kPa 8:01
VENTING 8:01
AIR DRYING 8:53
CYCLE COMPLETE 12:57

Digital Signature #
276DD8BE45A9D66F
