

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4395

Name: RITTER2  
Sterilizer ID: V2313228  
03/03/2025 12:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.2 F 0.0  
2:00 144.1 F 1.4  
4:00 145.0 F 4.0  
6:00 172.4 F 8.4  
8:00 197.7 F 7.7  
10:00 220.1 F 7.3  
12:00 234.1 F 12.6  
14:00 235.9 F 9.9  
16:00 235.9 F 8.3  
18:00 250.0 F 15.6  
20:00 262.7 F 22.9  
22:00 271.9 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.9  
1:00 272.6 F 29.1

C534: STERILIZE MODE  
STEAM TEMP LOW  
CHAMBER TEMP 270F  
CHAMBER PRESS 21.9PSI  
ELAPSED TIME 01:13  
\*\* ITEMS NOT STERILE \*\*

03/03/2025 12:39 PM