

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3724

Name: RITTER 1  
Sterilizer ID: V2178198  
09/06/2023 10:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.5 F 0.0  
2:00 193.5 F 2.8  
4:00 208.1 F 7.2  
6:00 219.3 F 12.6  
8:00 229.1 F 8.0  
10:00 239.6 F 13.5  
12:00 241.7 F 11.3  
14:00 239.1 F 9.9  
16:00 252.8 F 16.6  
18:00 264.4 F 23.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.1 F 28.6  
2:00 271.7 F 28.6  
3:00 272.0 F 28.4  
4:00 271.8 F 28.7  
  
Min 271.2 F 28.0  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.4 F 0.0  
5:00 230.1 F 0.0  
10:00 240.7 F 0.0  
15:00 232.3 F 0.0  
20:00 226.7 F 0.0  
25:00 228.7 F 0.0  
30:00 229.5 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 19:55  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:59:04  
CYCLE COMPLETE

09/06/2023 11:02 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_