

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001565  
11:01 11/05/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:01  
PRESSURIZING 5:37  
STERILIZING 6:57  
136.0 C 220kPa 6:57  
136.4 C 222kPa 7:27  
136.3 C 222kPa 7:57  
136.0 C 221kPa 8:27  
135.8 C 220kPa 8:57  
136.4 C 222kPa 9:27  
135.9 C 221kPa 9:57  
136.4 C 223kPa 10:27  
VENTING 10:27  
AIR DRYING 11:36  
CYCLE COMPLETE 41:36  
-----

Digital Signature #  
28EE6D17C1BAB461  
-----