
Midmark M11 - v1.0.0
Total Cycles: 5325

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2025 04:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.5 F 0.0
2:00 164.9 F 3.1
4:00 182.9 F 7.2
6:00 201.0 F 11.7
8:00 219.6 F 4.3
10:00 224.8 F 8.9
12:00 236.9 F 12.8
14:00 238.0 F 10.2
16:00 236.2 F 8.3
18:00 248.3 F 14.3
20:00 259.6 F 20.6
22:00 270.1 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 271.9 F 29.0
2:00 272.5 F 29.1
3:00 271.6 F 28.8
4:00 272.4 F 29.4

Min 270.8 F 28.2
Max 273.3 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.0 F 0.0
5:00 216.6 F 0.0
10:00 210.9 F 0.0
15:00 199.5 F 0.0
20:00 215.9 F 0.0
25:00 234.2 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 22:50
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:01:31
CYCLE COMPLETE

02/07/2025 05:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____