

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3581

Name: RITTER2  
Sterilizer ID: V2313228  
05/02/2024 09:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.8 F 0.0  
2:00 166.6 F 1.8  
4:00 173.2 F 4.7  
6:00 198.9 F 8.9  
8:00 212.2 F 13.2  
10:00 228.1 F 7.7  
12:00 237.3 F 12.1  
14:00 229.8 F 8.6  
16:00 240.0 F 12.0  
18:00 242.1 F 12.2  
20:00 253.6 F 18.0  
22:00 263.7 F 24.2  
24:00 272.2 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.6  
1:00 272.0 F 29.9  
2:00 271.8 F 30.1  
3:00 271.5 F 29.7  
4:00 272.2 F 30.3

Min 271.0 F 29.5  
Max 272.7 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.3 F 0.0  
5:00 237.2 F 0.0  
10:00 200.1 F 0.0  
15:00 180.1 F 0.0  
20:00 174.6 F 0.0  
25:00 172.8 F 0.0  
30:00 172.8 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 24:57  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:03:51  
CYCLE COMPLETE

05/02/2024 10:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_