
Midmark M11 - v1.0.5
Total Cycles: 254

Name: RITTER2
Sterilizer ID: V2313228
07/28/2021 06:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.3 F 0.0
2:00 80.6 F 0.0
4:00 122.8 F 0.1
6:00 115.5 F 0.1
8:00 194.0 F 2.4
10:00 170.8 F 6.6
12:00 185.8 F 10.5
14:00 225.7 F 6.6
16:00 233.8 F 9.2
18:00 242.1 F 12.5
20:00 239.3 F 10.6
22:00 234.4 F 8.2
24:00 247.6 F 14.4
26:00 258.3 F 20.5
28:00 268.6 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 272.9 F 29.8
1:00 272.2 F 30.0
2:00 272.0 F 29.9
3:00 272.0 F 29.9
4:00 272.0 F 30.1

Min 271.5 F 29.7
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.9 F 0.0
5:00 228.7 F 0.0
10:00 232.9 F 0.0
15:00 230.7 F 0.0
20:00 224.7 F 0.0
Dry Cycle Aborted

FILLING: 0:53
HEATING: 29:46
STERILIZING: 4:00
VENTING: 2:57
DRYING: 20:43
TOTAL CYCLE: 00:58:19

07/28/2021 07:22 AM
ACCEPTED: X
REJECTED:
Operator: U01