
Midmark M11 - v1.0.0
Total Cycles: 3315

Name: RITTER 1
Sterilizer ID: V2178198
05/31/2023 02:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.9 F 0.0
2:00 193.4 F 1.4
4:00 214.5 F 6.8
6:00 228.8 F 12.4
8:00 231.4 F 8.3
10:00 242.5 F 11.7
12:00 243.3 F 12.2
14:00 241.8 F 11.1
16:00 255.4 F 17.9
18:00 267.0 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.0
1:00 272.5 F 28.8
2:00 271.5 F 27.9
3:00 272.5 F 28.8
4:00 271.8 F 28.1

Min 271.4 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 228.6 F 0.0
10:00 241.1 F 0.0
15:00 230.2 F 0.0
20:00 226.4 F 0.0
25:00 228.8 F 0.0
30:00 228.1 F 0.0

Drying Complete

FILLING: 1:53
HEATING: 19:20
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

05/31/2023 02:58 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____