

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1819

Name: RITTER2  
Sterilizer ID: V2313228  
12/07/2022 02:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.1 F 0.0  
2:00 192.9 F 4.4  
4:00 200.9 F 9.7  
6:00 221.2 F 6.4  
8:00 233.0 F 11.7  
10:00 236.5 F 9.9  
12:00 236.4 F 8.9  
14:00 251.7 F 16.6  
16:00 264.4 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.6 F 29.5  
2:00 271.5 F 29.0  
3:00 272.7 F 29.7  
4:00 271.9 F 29.1  
  
Min 271.4 F 28.8  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.4 F 0.0  
5:00 232.9 F 0.0  
10:00 229.5 F 0.0  
15:00 227.7 F 0.0  
20:00 223.7 F 0.0  
25:00 220.8 F 0.0  
30:00 218.4 F 0.0  
Drying Complete

FILLING: 2:04  
HEATING: 17:49  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:56:58  
CYCLE COMPLETE

12/07/2022 03:54 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_