
Midmark M11 - v1.0.0
Total Cycles: 1629

Name: RITTER 1
Sterilizer ID: V2178198
02/14/2022 12:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.2 F 0.0
2:00 185.3 F 3.2
4:00 198.3 F 7.7
6:00 218.4 F 13.4
8:00 227.5 F 8.3
10:00 238.8 F 11.2
12:00 242.1 F 11.9
14:00 241.1 F 10.7
16:00 255.1 F 17.9
18:00 267.2 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.2 F 28.5
2:00 272.6 F 29.0
3:00 271.7 F 28.1
4:00 272.7 F 28.9

Min 271.5 F 28.1
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 224.8 F 0.0
10:00 240.1 F 0.0
15:00 229.1 F 0.0
20:00 225.0 F 0.0
25:00 224.9 F 0.0
30:00 224.9 F 0.0

Drying Complete

FILLING: 1:34
HEATING: 19:22
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

02/14/2022 01:36 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____