
Midmark M11 - v1.0.5
Total Cycles: 3354

Name: RITTER2
Sterilizer ID: V2313228
02/05/2024 10:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.2 F 0.0
2:00 151.0 F 0.7
4:00 154.2 F 3.2
6:00 176.4 F 6.6
8:00 193.9 F 10.9
10:00 215.8 F 4.7
12:00 226.5 F 9.2
14:00 231.9 F 10.9
16:00 237.8 F 11.1
18:00 236.3 F 9.4
20:00 250.0 F 15.9
22:00 261.5 F 22.3
24:00 271.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.5 F 30.0
2:00 272.7 F 29.6
3:00 271.6 F 29.2
4:00 271.2 F 29.2

Min 271.2 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 206.4 F 0.0
10:00 211.0 F 0.0
15:00 204.5 F 0.0
20:00 198.8 F 0.0
25:00 195.1 F 0.0
30:00 194.8 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:41
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:03:13
CYCLE COMPLETE

02/05/2024 11:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____