

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2536

Name: RITTER2  
Sterilizer ID: V2313228  
06/20/2023 09:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.0 F 0.0  
2:00 185.3 F 2.0  
4:00 184.2 F 5.9  
6:00 207.1 F 11.0  
8:00 219.7 F 6.4  
10:00 233.8 F 11.8  
12:00 233.5 F 9.1  
14:00 238.5 F 9.8  
16:00 246.9 F 14.2  
18:00 258.9 F 20.7  
20:00 269.3 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.3 F 29.3  
1:00 272.8 F 29.8  
2:00 272.0 F 29.8  
3:00 272.0 F 29.3  
4:00 272.4 F 29.9

Min 271.6 F 29.2  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.2 F 0.0  
5:00 207.7 F 0.0  
10:00 218.1 F 0.0  
15:00 218.3 F 0.0  
20:00 218.3 F 0.0  
25:00 220.0 F 0.0  
30:00 220.0 F 0.0  
Drying Complete

FILLING: 1:47  
HEATING: 21:36  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:00:38  
CYCLE COMPLETE

06/20/2023 10:39 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_