
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 006283
15:50 11/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 3:00
STERILIZING 4:04
136.5 C 224kPa 4:04
136.3 C 223kPa 4:34
136.4 C 223kPa 5:04
136.4 C 223kPa 5:34
136.3 C 223kPa 6:04
136.3 C 224kPa 6:34
136.3 C 222kPa 7:04
136.4 C 225kPa 7:34
136.4 C 225kPa 7:34
VENTING 7:34
AIR DRYING 8:27
CASSETTE REMOVED
CYCLE COMPLETE 14:47

Digital Signature #
69EA4FC52D4C446C
