
Midmark M11 - v1.0.5
Total Cycles: 183

Name: RITTER2
Sterilizer ID: V2313228
07/08/2021 12:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.7 F 0.0
2:00 179.3 F 3.9
4:00 203.2 F 9.8
6:00 221.9 F 6.3
8:00 234.4 F 13.4
10:00 240.0 F 12.0
12:00 241.8 F 11.5
14:00 256.1 F 19.0
16:00 268.1 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.9
1:00 271.5 F 29.0
2:00 271.6 F 29.1
3:00 272.1 F 29.3
4:00 272.3 F 29.6

Min 271.2 F 28.7
Max 272.4 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.9 F 0.0
5:00 234.0 F 0.0
10:00 216.1 F 0.0
15:00 213.7 F 0.0
20:00 209.8 F 0.0
25:00 209.8 F 0.0
30:00 209.6 F 0.0
Drying Complete

FILLING: 3:03
HEATING: 16:52
STERILIZING: 4:00
VENTING: 2:52
DRYING: 30:00
TOTAL CYCLE: 00:56:47
CYCLE COMPLETE

07/08/2021 01:09 PM
ACCEPTED: X
REJECTED:
Operator: U02