
Midmark M11 - v1.0.0
Total Cycles: 873

Name: RITTER 1
Sterilizer ID: V2178198
06/08/2021 12:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.0 F 0.0
2:00 157.2 F 1.9
4:00 172.2 F 5.0
6:00 196.8 F 9.1
8:00 220.7 F 9.5
10:00 227.6 F 9.6
12:00 233.0 F 8.7
14:00 238.4 F 11.9
16:00 238.3 F 10.0
18:00 250.7 F 16.4
20:00 262.2 F 23.0
22:00 272.7 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.0 F 29.5
2:00 272.1 F 29.7
3:00 272.3 F 29.8
4:00 272.2 F 29.8

Min 271.7 F 29.2
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 231.1 F 0.0
10:00 227.9 F 0.0
15:00 220.0 F 0.0
20:00 219.3 F 0.0
25:00 221.5 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:53
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:01:19
CYCLE COMPLETE

06/08/2021 01:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____