
Midmark M11 - v1.0.0
Total Cycles: 5034

Name: RITTER 1
Sterilizer ID: V2178198
10/29/2024 02:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.8 F 0.0
2:00 94.8 F 0.0
4:00 116.9 F 0.2
6:00 155.5 F 0.6
8:00 209.7 F 1.8
10:00 207.3 F 5.0
12:00 214.7 F 8.0
14:00 224.5 F 11.3
16:00 224.2 F 5.2
18:00 233.6 F 9.3
20:00 242.4 F 11.7
22:00 242.2 F 11.2
24:00 238.0 F 9.2
26:00 250.9 F 15.3
28:00 261.5 F 21.3
30:00 271.1 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.4 F 28.7
2:00 272.1 F 28.3
3:00 271.4 F 27.9
4:00 272.0 F 28.5

Min 271.2 F 27.8
Max 272.5 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.6 F 0.0
5:00 230.6 F 0.0
10:00 195.8 F 0.0
15:00 218.5 F 0.0
20:00 188.7 F 0.0
25:00 160.4 F 0.0
30:00 161.9 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 30:49
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:09:26
CYCLE COMPLETE

10/29/2024 03:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____