
Midmark M11 - v1.0.0
Total Cycles: 4485

Name: RITTER 1
Sterilizer ID: V2178198
04/24/2024 11:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.6 F 0.0
2:00 140.4 F 1.9
4:00 145.9 F 4.8
6:00 179.0 F 8.0
8:00 191.9 F 12.3
10:00 214.0 F 5.3
12:00 222.3 F 9.7
14:00 233.5 F 10.0
16:00 237.0 F 10.6
18:00 234.6 F 8.2
20:00 247.2 F 14.1
22:00 258.0 F 20.0
24:00 268.1 F 26.3
26:00 271.9 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.7 F 29.4
2:00 271.7 F 29.2
3:00 272.5 F 29.3
4:00 271.6 F 29.2

Min 271.5 F 28.9
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 209.9 F 0.0
10:00 226.8 F 0.0
15:00 232.1 F 0.0
20:00 228.5 F 0.0
25:00 230.7 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 26:06
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:04:47
CYCLE COMPLETE

04/24/2024 12:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____