

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2699

Name: RITTER2  
Sterilizer ID: V2313228  
07/25/2023 02:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.9 F 0.0  
2:00 165.9 F 0.7  
4:00 181.7 F 4.2  
6:00 195.9 F 8.0  
8:00 211.8 F 12.0  
10:00 221.4 F 6.1  
12:00 231.0 F 10.5  
14:00 233.3 F 8.5  
16:00 244.1 F 13.3  
18:00 243.8 F 12.5  
20:00 255.5 F 18.6  
22:00 266.0 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.4 F 29.4  
2:00 271.4 F 28.9  
3:00 272.3 F 29.4  
4:00 271.4 F 28.8

Min 271.3 F 28.7  
Max 272.5 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.6 F 0.0  
5:00 210.1 F 0.0  
10:00 209.8 F 0.0  
15:00 207.1 F 0.0  
20:00 204.2 F 0.0  
25:00 201.9 F 0.0  
30:00 199.7 F 0.0  
Drying Complete

FILLING: 1:34  
HEATING: 23:45  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:02:54  
CYCLE COMPLETE

07/25/2023 03:54 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_