
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 001494
10:40 10/13/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 4:56
STERILIZING 6:30
136.0 C 220kPa 6:30
136.4 C 218kPa 7:00
136.5 C 222kPa 7:30
135.9 C 219kPa 8:00
136.4 C 222kPa 8:30
136.4 C 222kPa 9:00
136.3 C 221kPa 9:30
136.0 C 218kPa 10:00
VENTING 10:00
AIR DRYING 11:01
CYCLE COMPLETE 41:01

Digital Signature #
3D4DB8DA8A8C6FD5
