
Midmark M11 - v1.0.0
Total Cycles: 1415

Name: RITTER 1
Sterilizer ID: V2178198
12/01/2021 11:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.0 F 0.0
2:00 182.8 F 3.7
4:00 198.9 F 8.3
6:00 220.4 F 8.6
8:00 232.8 F 9.8
10:00 234.8 F 8.3
12:00 244.2 F 13.3
14:00 243.0 F 12.2
16:00 256.2 F 19.1
18:00 267.9 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.7 F 29.5
2:00 271.9 F 28.8
3:00 272.2 F 29.4
4:00 272.8 F 29.4

Min 271.5 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 213.4 F 0.0
10:00 203.2 F 0.0
15:00 203.3 F 0.0
20:00 201.8 F 0.0
25:00 200.7 F 0.0
30:00 199.0 F 0.0
Drying Complete

FILLING: 2:11
HEATING: 19:18
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

12/01/2021 12:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____