
Midmark M11 - v1.0.5
Total Cycles: 722

Name: RITTER2
Sterilizer ID: V2313228
01/24/2022 11:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.1 F 0.0
2:00 152.5 F 3.1
4:00 176.5 F 7.3
6:00 199.7 F 12.3
8:00 217.6 F 6.4
10:00 231.0 F 11.9
12:00 234.2 F 9.5
14:00 237.2 F 9.3
16:00 248.3 F 15.1
18:00 260.7 F 22.0
20:00 271.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.7 F 29.8
2:00 271.7 F 29.1
3:00 271.9 F 29.7
4:00 272.6 F 29.7

Min 271.3 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.0 F 0.0
5:00 199.6 F 0.0
10:00 204.4 F 0.0
15:00 203.7 F 0.0
20:00 198.0 F 0.0
25:00 196.7 F 0.0
30:00 202.9 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 20:36
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:07
CYCLE COMPLETE

01/24/2022 12:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____