
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001119
17:35 09/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 3:52
STERILIZING 5:14
136.0 C 220kPa 5:14
136.4 C 226kPa 5:44
136.4 C 225kPa 6:14
136.3 C 223kPa 6:44
136.3 C 224kPa 7:14
136.4 C 223kPa 7:44
136.4 C 225kPa 8:14
136.2 C 224kPa 8:44
136.2 C 224kPa 8:44
VENTING 8:44
AIR DRYING 9:42
CYCLE COMPLETE 24:43

Digital Signature #
C3AC06043F814BDA
