
Midmark M11 - v1.0.0
Total Cycles: 924

Name: RITTER 1
Sterilizer ID: V2178198
06/22/2021 10:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.8 F 0.0
2:00 150.0 F 0.5
4:00 163.8 F 2.9
6:00 186.7 F 7.0
8:00 204.9 F 12.2
10:00 220.2 F 6.5
12:00 234.0 F 11.8
14:00 235.1 F 8.6
16:00 238.9 F 10.0
18:00 245.1 F 13.0
20:00 257.7 F 19.8
22:00 269.0 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 271.5 F 28.9
2:00 271.6 F 29.1
3:00 272.4 F 29.6
4:00 272.8 F 29.5

Min 270.9 F 28.8
Max 273.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 162.7 F 0.0
5:00 226.1 F 0.0
10:00 191.8 F 0.0
15:00 183.8 F 0.0
20:00 181.9 F 0.0
25:00 181.9 F 0.0
30:00 182.7 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 23:02
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:01:07
CYCLE COMPLETE

06/22/2021 11:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____