
Midmark M11 - v1.0.5
Total Cycles: 3377

Name: RITTER2
Sterilizer ID: V2313228
02/12/2024 12:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.7 F 0.0
2:00 154.9 F 2.3
4:00 166.7 F 5.7
6:00 195.8 F 10.3
8:00 214.5 F 5.1
10:00 226.8 F 9.2
12:00 234.7 F 10.1
14:00 240.7 F 12.0
16:00 239.8 F 10.8
18:00 253.2 F 17.4
20:00 264.8 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.5 F 29.5
2:00 271.4 F 29.1
3:00 272.6 F 29.6
4:00 271.5 F 28.9

Min 271.3 F 28.9
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.0 F 0.0
5:00 229.5 F 0.0
10:00 201.8 F 0.0
15:00 198.8 F 0.0
20:00 197.2 F 0.0
25:00 197.3 F 0.0
30:00 206.2 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:51
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

02/12/2024 01:22 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____