
Midmark M11 - v1.0.5
Total Cycles: 4039

Name: RITTER2
Sterilizer ID: V2313228
10/17/2024 01:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.3 F 0.0
2:00 124.6 F 0.0
4:00 134.6 F 1.3
6:00 168.9 F 4.7
8:00 188.4 F 8.4
10:00 201.3 F 12.9
12:00 219.1 F 7.3
14:00 227.9 F 12.1
16:00 232.5 F 9.0
18:00 240.5 F 10.6
20:00 245.2 F 13.4
22:00 256.9 F 19.7
24:00 267.6 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 271.6 F 29.6
2:00 271.7 F 29.8
3:00 271.7 F 29.8
4:00 271.7 F 29.8

Min 271.4 F 29.2
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 209.1 F 0.0
10:00 191.2 F 0.0
15:00 185.4 F 0.0
20:00 181.6 F 0.0
25:00 177.6 F 0.0
30:00 176.4 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:15
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:03:47
CYCLE COMPLETE

10/17/2024 02:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____