
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004788
13:02 29/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:55
STERILIZING 4:05
136.4 C 220kPa 4:05
136.3 C 224kPa 4:35
136.3 C 223kPa 5:05
136.3 C 224kPa 5:35
136.4 C 225kPa 6:05
136.3 C 224kPa 6:35
136.1 C 220kPa 7:05
136.3 C 223kPa 7:35
136.3 C 223kPa 7:35
VENTING 7:35
AIR DRYING 8:35
CYCLE COMPLETE 8:40

Digital Signature #
A156AC921CB0D40E
