

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1181

Name: RITTER2  
Sterilizer ID: V2313228  
06/09/2022 10:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.1 F 0.0  
2:00 185.6 F 3.5  
4:00 194.8 F 8.7  
6:00 224.4 F 8.5  
8:00 233.7 F 10.5  
10:00 233.4 F 8.3  
12:00 241.2 F 10.8  
14:00 246.3 F 14.0  
16:00 258.8 F 20.8  
18:00 269.9 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.6  
1:00 272.3 F 29.8  
2:00 272.7 F 30.2  
3:00 271.8 F 29.5  
4:00 272.4 F 30.1  
  
Min 271.7 F 29.5  
Max 272.8 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.5 F 0.0  
5:00 231.6 F 0.0  
10:00 228.1 F 0.0  
15:00 231.1 F 0.0  
20:00 228.9 F 0.0  
25:00 227.5 F 0.0  
30:00 229.1 F 0.0  
  
Drying Complete

FILLING: 2:15  
HEATING: 19:39  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:05  
CYCLE COMPLETE

06/09/2022 11:27 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_