
Midmark M11 - v1.0.0
Total Cycles: 2680

Name: RITTER 1
Sterilizer ID: V2178198
12/15/2022 08:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.2 F 0.0
2:00 88.0 F 0.0
4:00 122.0 F 0.0
6:00 158.9 F 0.0
8:00 184.4 F 1.1
10:00 202.3 F 4.0
12:00 214.8 F 7.6
14:00 225.8 F 11.6
16:00 228.2 F 5.0
18:00 231.0 F 9.3
20:00 241.4 F 11.5
22:00 239.4 F 10.9
24:00 236.9 F 8.6
26:00 250.2 F 15.3
28:00 261.7 F 21.8
30:00 272.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.3
1:00 272.8 F 29.1
2:00 271.7 F 28.1
3:00 272.8 F 29.0
4:00 271.7 F 28.2

Min 271.6 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.3 F 0.0
5:00 204.8 F 0.0
10:00 185.5 F 0.0
15:00 186.7 F 0.0
20:00 184.5 F 0.0
25:00 181.3 F 0.0
30:00 180.6 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 30:17
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:23
CYCLE COMPLETE

12/15/2022 09:09 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____