
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001952
13:23 14/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:52
STERILIZING 4:03
136.3 C 223kPa 4:03
136.2 C 223kPa 4:33
136.3 C 220kPa 5:03
136.3 C 224kPa 5:33
136.4 C 224kPa 6:03
136.2 C 223kPa 6:33
136.4 C 223kPa 7:03
136.4 C 225kPa 7:33
136.4 C 225kPa 7:33
VENTING 7:33
AIR DRYING 8:24
CYCLE COMPLETE 33:51

Digital Signature #
26350A500A6CE34D
