
Midmark M11 - v1.0.5
Total Cycles: 1420

Name: RITTER2
Sterilizer ID: V2313228
08/25/2022 04:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.1 F 0.0
2:00 194.1 F 3.2
4:00 206.1 F 8.5
6:00 231.5 F 10.0
8:00 234.3 F 11.1
10:00 235.3 F 9.3
12:00 237.9 F 9.3
14:00 250.1 F 15.6
16:00 262.1 F 22.8
18:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.3 F 29.5
2:00 271.4 F 29.2
3:00 272.3 F 29.6
4:00 271.5 F 28.9

Min 271.3 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 229.3 F 0.0
10:00 243.0 F 0.0
Dry Cycle Aborted

FILLING: 2:08
HEATING: 18:15
STERILIZING: 4:00
VENTING: 3:07
DRYING: 13:05
TOTAL CYCLE: 00:40:34

08/25/2022 05:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____