
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000791
13:23 15/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 6:04
PRESSURIZING 8:22
STERILIZING 10:39
136.0 C 220kPa 10:39
136.4 C 222kPa 11:09
136.1 C 217kPa 11:39
136.1 C 221kPa 12:09
136.4 C 223kPa 12:39
136.2 C 220kPa 13:09
136.4 C 223kPa 13:39
136.4 C 222kPa 14:09
136.4 C 222kPa 14:09
VENTING 14:09
AIR DRYING 15:24
CYCLE COMPLETE 45:24

Digital Signature #
2A3C0F6040886902
