

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 006778  
10:54 21/12/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:14  
PRESSURIZING 3:16  
STERILIZING 4:22  
136.4 C 220kPa 4:22  
136.1 C 217kPa 4:52  
136.4 C 224kPa 5:22  
136.4 C 223kPa 5:52  
136.3 C 221kPa 6:22  
136.3 C 220kPa 6:52  
136.4 C 223kPa 7:22  
136.2 C 217kPa 7:52  
136.2 C 217kPa 7:52  
VENTING 7:52  
AIR DRYING 8:44  
CYCLE COMPLETE 8:48

-----  
Digital Signature #  
B748FA4B09AD0DD6  
-----