

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2906

Name: RITTER2  
Sterilizer ID: V2313228  
09/19/2023 02:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.3 F 0.0  
2:00 155.8 F 0.2  
4:00 182.6 F 4.1  
6:00 204.3 F 9.1  
8:00 225.8 F 9.3  
10:00 233.7 F 9.5  
12:00 235.6 F 9.3  
14:00 244.1 F 13.0  
16:00 243.2 F 12.1  
18:00 256.1 F 19.0  
20:00 267.7 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.3 F 29.1  
2:00 272.2 F 29.4  
3:00 271.9 F 28.8  
4:00 272.6 F 29.6

Min 271.5 F 28.6  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.9 F 0.0  
5:00 232.5 F 0.0  
10:00 223.5 F 0.0  
15:00 196.2 F 0.0  
20:00 189.0 F 0.0  
25:00 186.5 F 0.0  
30:00 187.2 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 21:18  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 00:59:49  
CYCLE COMPLETE

09/19/2023 03:16 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_