
Midmark M11 - v1.0.5
Total Cycles: 3756

Name: RITTER2
Sterilizer ID: V2313228
07/05/2024 11:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.6 F 0.0
2:00 133.1 F 1.5
4:00 140.8 F 4.0
6:00 177.1 F 7.6
8:00 198.0 F 11.7
10:00 212.9 F 5.4
12:00 224.2 F 9.6
14:00 232.3 F 10.2
16:00 236.4 F 10.9
18:00 234.8 F 8.8
20:00 248.0 F 15.0
22:00 259.1 F 21.1
24:00 268.9 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.5
1:00 272.6 F 29.9
2:00 271.2 F 29.7
3:00 273.2 F 30.8
4:00 272.8 F 30.5

Min 270.9 F 29.3
Max 273.8 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 204.5 F 0.0
10:00 205.3 F 0.0
15:00 198.0 F 0.0
20:00 198.4 F 0.0
25:00 212.2 F 0.0
30:00 210.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 25:07
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:03:47
CYCLE COMPLETE

07/05/2024 12:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____