

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000520  
11:36 28/05/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:55  
PRESSURIZING 2:59  
STERILIZING 4:16  
136.0 C 227kPa 4:16  
136.4 C 233kPa 4:46  
136.1 C 230kPa 5:16  
136.4 C 233kPa 5:46  
136.4 C 232kPa 6:16  
136.5 C 233kPa 6:46  
136.4 C 232kPa 7:16  
136.2 C 231kPa 7:46  
136.2 C 231kPa 7:46  
VENTING 7:46  
AIR DRYING 8:42  
CYCLE COMPLETE 12:55  
-----

Digital Signature #  
BFEF5775ACAA81E9  
-----