

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001561  
9:02 11/05/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:20  
PRESSURIZING 7:08  
STERILIZING 8:47  
136.0 C 220kPa 8:47  
136.0 C 217kPa 9:17  
136.3 C 221kPa 9:47  
136.0 C 221kPa 10:18  
136.4 C 219kPa 10:48  
136.4 C 222kPa 11:18  
136.1 C 221kPa 11:48  
136.4 C 223kPa 12:18  
136.4 C 223kPa 12:18  
VENTING 12:18  
AIR DRYING 13:22  
CYCLE COMPLETE 43:04

-----  
Digital Signature #  
12898C58B5FB21C4  
-----