
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002456
8:31 18/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:51
PRESSURIZING 3:51
STERILIZING 4:58
136.4 C 221kPa 4:58
136.3 C 224kPa 5:28
136.2 C 220kPa 5:58
136.0 C 221kPa 6:28
136.4 C 223kPa 6:58
136.4 C 224kPa 7:28
136.2 C 222kPa 7:58
136.4 C 222kPa 8:28
136.4 C 222kPa 8:28
VENTING 8:28
AIR DRYING 9:24
CYCLE COMPLETE 9:26

Digital Signature #
159F1B33224B500C
