
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001946
14:26 06/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:33
PRESSURIZING 3:33
STERILIZING 4:31
136.0 C 220kPa 4:31
136.0 C 221kPa 5:01
136.2 C 224kPa 5:31
136.4 C 223kPa 6:01
136.1 C 219kPa 6:31
136.2 C 223kPa 7:01
136.3 C 224kPa 7:31
136.5 C 225kPa 8:01
136.5 C 225kPa 8:01
VENTING 8:01
AIR DRYING 8:52
CYCLE COMPLETE 8:59

Digital Signature #
78A64B2AFEC302BC
