
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000185
12:59 06/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:04
PRESSURIZING 4:05
STERILIZING 5:06
136.0 C 221kPa 5:06
136.4 C 225kPa 5:36
136.4 C 224kPa 6:06
136.3 C 224kPa 6:36
136.4 C 224kPa 7:06
136.4 C 224kPa 7:36
136.3 C 224kPa 8:06
136.2 C 220kPa 8:36
136.2 C 220kPa 8:36
VENTING 8:36
AIR DRYING 9:32
CYCLE COMPLETE 9:34

Digital Signature #
BAB0B8FBF5F7F036
