
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001726
13:56 27/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:05
STERILIZING 3:06
136.0 C 222kPa 3:06
136.3 C 224kPa 3:36
136.4 C 225kPa 4:06
136.3 C 224kPa 4:36
136.2 C 224kPa 5:06
136.4 C 225kPa 5:36
136.4 C 226kPa 6:06
136.1 C 223kPa 6:36
136.1 C 223kPa 6:36
VENTING 6:36
AIR DRYING 7:29
CYCLE COMPLETE 7:32

Digital Signature #
3AB3623B938CB66A
