
Midmark M11 - v1.0.5
Total Cycles: 646

Name: RITTER2
Sterilizer ID: V2313228
12/16/2021 02:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.7 F 0.0
2:00 177.4 F 4.3
4:00 199.1 F 10.3
6:00 218.1 F 6.0
8:00 234.0 F 12.5
10:00 238.0 F 11.2
12:00 238.8 F 10.8
14:00 253.6 F 18.1
16:00 265.8 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.2 F 29.6
2:00 272.6 F 30.2
3:00 271.8 F 29.2
4:00 272.7 F 30.1

Min 271.6 F 29.2
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 201.1 F 0.0
10:00 188.0 F 0.0
15:00 192.1 F 0.0
20:00 191.7 F 0.0
25:00 266.6 F 0.0
Dry Cycle Aborted

FILLING: 2:25
HEATING: 17:34
STERILIZING: 4:00
VENTING: 3:08
DRYING: 29:44
TOTAL CYCLE: 00:56:51

12/16/2021 03:37 PM
ACCEPTED: X
REJECTED:
Operator: U04