
Midmark M11 - v1.0.5
Total Cycles: 3202

Name: RITTER2
Sterilizer ID: V2313228
12/11/2023 10:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.1 F 0.0
2:00 142.0 F 0.8
4:00 156.7 F 3.2
6:00 181.0 F 6.9
8:00 201.9 F 11.2
10:00 217.8 F 5.0
12:00 226.1 F 9.9
14:00 233.2 F 9.0
16:00 241.4 F 12.2
18:00 239.6 F 10.7
20:00 252.3 F 16.9
22:00 262.9 F 23.3
24:00 272.7 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.4 F 29.5
2:00 272.6 F 29.9
3:00 271.5 F 29.7
4:00 272.1 F 29.5

Min 271.3 F 29.3
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.8 F 0.0
5:00 236.2 F 0.0
10:00 189.1 F 0.0
15:00 188.0 F 0.0
20:00 184.9 F 0.0
25:00 182.1 F 0.0
30:00 180.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:04
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:03:27
CYCLE COMPLETE

12/11/2023 11:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____