

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3326

Name: RITTER 1  
Sterilizer ID: V2178198  
06/02/2023 09:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 132.7 F 0.0  
2:00 134.8 F 0.0  
4:00 138.4 F 0.0  
6:00 168.9 F 1.8  
8:00 191.5 F 5.0  
10:00 204.3 F 8.9  
12:00 218.5 F 13.1  
14:00 225.2 F 7.1  
16:00 233.5 F 11.7  
18:00 233.9 F 8.3  
20:00 244.5 F 12.6  
22:00 243.9 F 12.1  
24:00 255.2 F 17.9  
26:00 265.1 F 23.9  
28:00 271.3 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.4 F 29.0  
2:00 272.2 F 28.7  
3:00 271.2 F 28.1  
4:00 271.7 F 28.7  
  
Min 271.0 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.5 F 0.0  
5:00 208.5 F 0.0  
10:00 222.6 F 0.0  
15:00 232.5 F 0.0  
20:00 227.1 F 0.0  
25:00 225.3 F 0.0  
30:00 226.1 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 28:06  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:06:32  
CYCLE COMPLETE

06/02/2023 10:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_